
Instructions for use / installation of the boso profil-manager XD Home

**Powerful software for computer-assisted
blood pressure control and observation of therapy**

Compatible with the following blood pressure meters:

**medilife PC2
medilife PC3**

These instructions for use and installation of the boso profil-manager XD Home apply to version 4.2.0.133 and above. Please refer to the readme.txt file on the installation CD for new information.

The drive letters in these instructions for use are simply examples:

C:\ for the local hard drive

D:\ for the CD/DVD drive

The path data corresponds to a standard single-user installation under the german Windows 7 (32-bit) operating system and may be different in other operating systems or language versions.

Table of contents

1.1	Before installation.....	6
1.1.1	Data backup.....	6
1.1.2	Operating the software	6
1.2	Installing boso profil-manager XD Home.....	7
1.3	Converting data from profilmanager 3.....	10
1.4	Installing the USB connection cable.....	12
2.	Start.....	13
2.1	Patient registry, for all applications.....	14
2.1.1	„New...“ button	14
2.1.2	„Modify...“ button	14
2.1.3	„Delete...“ button	14
2.1.4	„Export/Import“ button.....	14
2.1.4.1	„Export Excel“	14
2.1.4.2	„Export patient (XML)“	14
2.1.4.3	„Import patient (XML)“	15
2.1.5	„Import GDT“	15
2.1.6	„Settings...“ button	16
2.1.7	„Print...“ button	16
3.	„Patient measurement“ application.....	17
3.1	„Patient“ registry	17
3.2	„Measurements“ registr	17
3.2.1	„Delete period“ button	19
3.2.2	„Export GDT“ button	19
3.2.3	„Manual data“ button	19
3.2.4	„Print...“ button.....	19
3.2.5	„Trasfer data“ button.....	20
3.2.6	„Interval limits“ button.....	21
3.2.7	„All, Morning, Afternoon“ selection.....	21
3.2.8	„Period“ selection	21
3.3	„Sys/dia profile“ registry.....	22
3.3.1	„All, Morning, Afternoon“ selection.....	22
3.3.2	„Period“ selection	23
3.3.3	„Time scope“ selection field	23
3.3.4	„Print...“ button.....	23
3.4	„MAP profile“ registry.....	24
3.4.1	„All, Morning, Afternoon“ selection.....	24
3.4.2	„Period“ selection	25
3.4.3	„Time scope“ selection field	25
3.4.4	Print...“ button.....	25
3.5	„PP profile“ registry.....	26
3.5.1	„All, Morning, Afternoon“ selection.....	26
3.5.2	„Time scope“ selection field	27
3.5.3	„Period“ selection	27

3.5.4	„Print...” button.....	27
3.6	„Statistics” button.....	28
3.6.1	„Print...” button.....	29
3.7	„Settings...” button in „Patient” registry	30
3.7.1	„Doctor” sub-registry	30
3.7.2	„Import/export” sub-registry.....	31
3.7.3	„Printer settings” sub-registry.....	32
3.7.4	„Interface” sub-registry	33
3.7.5	„Limits” sub-registry.....	34
3.7.6	„Profile” sub-registry.....	35
4.	Warranty provisions.....	36
5.	Limitation of liability.....	37
6.	Explanation of icons.....	37
Appendix A – Extract from GDT record description		38

Contents of package

1 installation CD

1 user manual boso profil-manager XD Home

Minimum operating requirements:

- Pentium II or higher
- 512 MB RAM
- Graphics card
- 100 MB free hard disk space
- USB 1.1 interface
- Microsoft Windows® XP

Accuracy of the values displayed

Systole,
diastole,
puls = accuracy of the measuring instruments used

Intended use of a medical device

Software for

- representation of the determined values recorded by boso measuring instruments
- archiving of measured values in a separate database
- printout of the data for transmission to the doctor

1.1 Before installation...

You have chosen boso profil-manager XD Home, a software solution which gives you optimum results with minimum input time. It has been deliberately designed to be operated without the usual menu structure based on individual registries. This is why the user manual also refers to individual windows and program parts as registries.

1.1.1 Data backup

To avoid losing data, you should back up the 'profman.fdb' file from the C:\ProgramData\Boso\profmanXD\ sub-directory at regular intervals.

1.1.2 Operating the software

The program is operated by the mouse.



PC requirements

Computers used to assess data must comply with the latest version of European standard EN 60601-1.


1.2 Installing boso profil-manager XD Home

To install boso profil-manager XD Home, start file D:\ Setup_profmanXDHome.exe manually or automatically via the autostart function which runs when you have inserted the CD.




Make sure that you have administrator rights when installing the program (if in doubt, start the "Setup_profmanXDHome.exe" program with the right mouse button, "execute as administrator").

Select your preferred language for installation (screenshot in German):




The available languages are German, English and Hungarian.

You can change the target folder for installation individually, but in most cases the default folder is used unchanged:



At the end of installation, tick the “Install boso USB Driver” checkbox to install the USB driver needed for the PC connection cable.




Once you have successfully installed the single-user version of boso profil-manager XD Home, click on the “Close” button to close the installation screen.

1.3 Converting data from boso profil-manager 3

Use the Konverter.exe program to convert existing data from earlier Firebird databases.

Launch the program either via Windows Explorer from the directory path C:\Programme\Boso\ProfmanXD\ or via the Windows start button, selecting the program “Transfer data from boso profil-manager 3” under the “boso profil-manager XD Home” program group (file path in German)



The databases to be converted must be on the local system. Conversion via the network is not possible.

Enter a username and password to register with the Firebird database server. The default username is "sysdba" and the default password is "masterkey".

Select the storage location for the profil-manager XD Home file in the "Target DB (profil-manager XD Home)" field.

Select the storage location of the database you want to convert (profil-manager 3). Click on the "Connect" button to connect to the source database.

Once you have established a connection with the source database, start transferring data by clicking on the button of the same name ("Start transfer") (screenshot in German).

The screenshot shows a Windows-style dialog box titled "Datenübernahme ABI bzw. profilmanager 3 -> profilmanagerXD". It contains the following elements:

- 1. Server:** An empty text field.
- 2. Benutzername:** A text field containing "Sysdba".
- 3. Passwort:** A text field containing "****".
- 4. Ziel-DB (profilmanager XD):** A text field containing "C:\ProgramData\Boso\profmanXD\profman.fdb" with a browse button "...".
- Installationstyp: Lokal**
- Tabs:** Two tabs are present: "Datenübernahme aus ABI" and "Datenübernahme aus Profilmanager 3". The second tab is selected.
- Quell-DB (Profman 3):** A text field containing "C:\ProgramData\Boso\Profman\bosom_tm+pm.gdb" with a browse button "...".
- Buttons:** "Connect" and "Übernahme starten" are located below the source database field.
- Wichtiger Hinweis:** A section with the text: "Die Konvertierung muss auf dem Server durchgeführt werden. Alle Files müssen sich auf dem lokalen Filesystem (kein Netzwerklauferwerk) befinden."
- Beenden:** A button at the bottom left.
- Version 1.0.0.12** is displayed at the bottom left corner.

1.4 Installing the USB connection cable

If the USB driver was pre-installed during installation (this is normally the case), you simply need to plug the USB connection cable provided into a free USB port. The cable is then automatically recognised and installed.

If the driver was not pre-installed, you must execute the file `dpinst_32bit.exe` (present after installation in the directory `C:\Programme\Boso\profmanXD\ftdi\`) as administrator and then plug the USB cable into a free USB port.

2. Start

boso

boso profil-manager XD

Name:

Forename:

Patient no.:

Birthdate:

Patient measurement

boso

Patient

Measurements

Sys/dia profile

MAP profile

PP profile

Statistics

Search:

Patient no.	Surname ...	First name ...	Date of birth ...	Place (Postal code) ...	Remark
1-B	Mustermann	Manfred			

New...

Edit...

Delete...

Export/Import ▾

Import GDT

Settings...

Print...

Billing list

Version 4.2.0.132

CRC32: 7EB8D631


User: Sysdba

Server: local

Pattern: 50

DB: C:\ProgramData\Boso\profmanXD\...

Accessing the help function

Press the F1 key or  button to call up these and other instructions for use in PDF format at any time.

2.1 Patient registry

Click on a patient to select him or her.

Patients can be sorted by clicking on the header of the relevant column.

Use the “Restore column order” function to go back to the original sorting by patient number. To do this, right-click in the header row of the patient table.

You can search by name via the “Search” function.

2.1.1 “New...” button

This sets up a new patient. The name and patient number fields must be completed. The system will suggest a patient number, but you can change this manually.

2.1.1 “Modify...” button

This alters an existing patient’s personal data.

2.1.1 “Delete...” button

This deletes a patient.



This process is irreversible: all data (including measurements from all parts of the application) are deleted and cannot be recovered.

2.1.4 “Export/Import” button

2.1.4.1 “Export Excel” exports data for the currently selected patient in Excel-readable .xls format. You can choose whatever file name and storage location you like.

2.1.4.2 “Export patient (XML)” exports data for the currently selected patient in profil-manager XD Home-readable .xml format.

2.1.4.3 "Import Patient (XML)" imports data for a selected patient.


The following rules apply to 2.1.4.2 and 2.1.4.3:

The transfer path from the transfer file settings is used for incorporation into the clinic's data processing system (see next chapter, "Settings..." button)

The file name is made up of the active application and the patient number.
For example, "medicus_10.xml" stands for patient data for the patient with patient number 10.

2.1.5 Import GDT

The GDT-import file, if available, is read and the included patient is selected or created. If a patient with the imported patient number already exists, the following screen appears.



The dialog box titled "Do you want to select this patient?" contains the following fields and controls:

- Personal Data**
 - Name:
 - Forename:
 - Date of birth: (format: dd.mm.yyyy)
 - Gender: ☒ Male, ☐ Female
- Other**
 - Patient No.:
 - Insurance No.:
- Address**
 - Street:
 - Zip / City:
- Dimensions**
 - Height: cm
 - Weight: kg
- Remarks**
 -
- Buttons**
 - Yes, select this patient (highlighted in blue)
 - No, use patient data from GDT import
 - Cancel

The button "Yes, select this patient" selects the indicated patient.



With the button "No, use patient data from GDT import" all existing data will be assigned to the patient from the GDT import file.

2.1.6 “Settings” button

The settings are explained in the sections dealing with particular applications (chapters 3 to 5) as they are application-specific.

2.1.7 “Print...” button

Patient data is presented on screen (preview), printed out on the printer selected under “Setup”, and/or in the configured PDF file.

The Windows standard printer is the default for print-outs.

If you selected “PDF file” in the printer settings, a PDF file will be created. The name of this PDF file is pmXD_pat#.pdf

Where

pat#=patient number

3. "Patient measurement" application

3.1 "Patient" registry

see section "2.1 Patient registry, for all applications"

See chapter 3.7 for "Settings" button.

3.2 “Measurements” registry

boso

boso profil-manager XD

Name: **Mustermann** Forename: **Manfred**

Patient no.: **1-8** Birthdate:

Patient measurement

Patient		Measurements		Sys/dia profile		MAP profile		PP profile		Statistics	
Date	SYS	DIA	PUL	MAP	PP	Arr.	Exclude	Remark	Period	Device	
1 Mi, 06.04.2005 - 18:26	144	75	65	98	69		<input type="checkbox"/>		06.04.2005 - 27.06.2005 medlife PC3 <input type="button" value="Delete Period..."/> <input type="button" value="Export GDT"/> <input type="button" value="Manual Data..."/> <input type="button" value="Print..."/> <input type="button" value="Transfer Data..."/> <input type="button" value="Interval limits..."/>		
2 Do, 07.04.2005 - 08:32	159	88	68	112	71		<input type="checkbox"/>				
3 Do, 07.04.2005 - 18:37	153	92	67	112	61		<input type="checkbox"/>				
4 Fr, 08.04.2005 - 08:44	164	76	72	105	88		<input type="checkbox"/>				
5 Fr, 08.04.2005 - 18:49	152	87	68	109	65		<input type="checkbox"/>				
6 Sa, 09.04.2005 - 08:54	176	87	67	117	89		<input type="checkbox"/>				
7 Sa, 09.04.2005 - 18:59	160	99	67	119	61		<input type="checkbox"/>				
8 So, 10.04.2005 - 08:04	146	80	68	102	66		<input type="checkbox"/>				
9 So, 10.04.2005 - 18:09	153	82	66	106	71		<input type="checkbox"/>				
10 Mo, 11.04.2005 - 08:16	156	91	61	113	65		<input type="checkbox"/>				
11 Mo, 11.04.2005 - 18:20	160	84	65	109	76		<input type="checkbox"/>				
12 Di, 12.04.2005 - 08:28	151	88	60	109	63		<input type="checkbox"/>				
13 Di, 12.04.2005 - 18:34	152	81	64	105	71		<input type="checkbox"/>				
14 Mi, 13.04.2005 - 08:39	152	81	65	105	71		<input type="checkbox"/>				
15 Mi, 13.04.2005 - 18:46	177	86	62	116	91		<input type="checkbox"/>				
16 Do, 14.04.2005 - 08:51	161	89	60	113	72		<input type="checkbox"/>				
17 Do, 14.04.2005 - 18:57	163	86	70	112	77		<input type="checkbox"/>				
18 Fr, 15.04.2005 - 08:02	174	92	67	119	82		<input type="checkbox"/>				
19 Fr, 15.04.2005 - 18:07	166	96	71	119	70		<input type="checkbox"/>				
20 Sa, 16.04.2005 - 08:12	164	91	64	115	73		<input type="checkbox"/>				
21 Sa, 16.04.2005 - 18:17	148	91	70	110	57		<input type="checkbox"/>				
22 So, 17.04.2005 - 08:22	138	91	70	107	47		<input type="checkbox"/>				

☒ All
 ☐ Morning
 ☐ Afternoon
 ☐ Period
 06.04.2005 - 27.06.2005

Version 4.2.0.132 CRC32: 7EB8D631 User: Sysdba Server: local Pattern: 50 DB: C:\ProgramData\Boso\profmangXD\

All measurements that have been read in to date for the selected patient are displayed here. Both the period and the blood pressure meter with which the series of measurements was carried out are displayed.

	Date	SYS	DIA	PUL	MAP	PP	Arr.	Exclude	Remark
1	Mi, 06.04.2005 - 18:26	144	75	65	98	69		<input type="checkbox"/>	Test
2	Do, 07.04.2005 - 08:32	159	88	68	112	71		<input type="checkbox"/>	
3	Do, 07.04.2005 - 18:37	153	92	67	112	61		<input type="checkbox"/>	
4	Fr, 08.04.2005 - 08:44	164	76	72	105	88		<input type="checkbox"/>	
5	Fr, 08.04.2005 - 18:49	152	87	68	109	65		<input type="checkbox"/>	
6	Sa, 09.04.2005 - 08:54	176	87	67	117	89		<input type="checkbox"/>	
7	Sa, 09.04.2005 - 18:59	160	99	67	119	61		<input type="checkbox"/>	
8	So, 10.04.2005 - 08:04	146	80	68	102	66		<input type="checkbox"/>	
9	So, 10.04.2005 - 18:09	153	82	66	106	71		<input type="checkbox"/>	
10	Mo, 11.04.2005 - 08:16	156	91	61	113	65		<input type="checkbox"/>	
11	Mo, 11.04.2005 - 18:20	160	84	65	109	76		<input type="checkbox"/>	
12	Di, 12.04.2005 - 08:28	151	88	60	109	63		<input type="checkbox"/>	
13	Di, 12.04.2005 - 18:34	152	81	64	105	71		<input type="checkbox"/>	
1	2	3	4	5	6	7	8	9	10

The column content is as follows:

1: serial numbers of measurements

2: date and time of measurement

3: SYS = systolic blood pressure

4: DIA = diastolic blood pressure

5: PUL = pulse

6: MAP = mean arterial blood pressure, calculated as $DIA + 1/3 \text{ pulse pressure}$

7: PP = pulse pressure (SYS – DIA)

8: Arr. = indication of arrhythmia

9: Exclude = if you highlight this, the measurement will not be taken into account for assessment in the profile and statistics

10: Remark. Double-click in the measurement line to add a remark to a measurement.

This opens the remark input field:

Process remark

Date06.04.2005Time18:26

Sys144

Dia75

Puls65

MAP98

PP69

Remark

Test

OK

Cancel

3.2.1 “Delete period” button

This irrevocably deletes a complete set of readings.

3.2.2 “Export GDT” button

This generates the GDT export file manually. Only available if the GDT import of the active patient has already been performed.

3.2.3 “Manual data” button

Additional readings can be added to the active set of readings here.

3.2.4 “Print...” button

This prints out the list of readings.

If you selected “PDF file” in the printer settings (chapter 3.7.3), a PDF file will be created. The name of this PDF file is `medi_pat#_YYYYMMDD_m.pdf`

where


pat#=patient number

YYYYMMDD = date of measurement


m = measurement reference letter

3.2.5 "Transfer data" button


Connect the meter to the PC.



Once you have clicked on the "Transfer data" button, the START/STOP button on the blood pressure meter must be pressed within 6 seconds in order to transfer data from the blood pressure meter. Measurements will be transferred from the meter which is connected.



Once data has been successfully transferred, you can set up a new period or assign the data to an existing period.



3.2.6 “Interval limits” button

The limits for morning and afternoon measurements can be defined here. The default setting is midnight to 11.59 a.m. for morning and noon to 11.59 p.m. for afternoon.

3.2.7 “All, Morning, Afternoon” selection

You can restrict the reading selection to only morning measurements or only afternoon measurements here.

3.2.8 “Period” selection

☒ Period 06.04.2005 - 10.02.2012

You can select a specific period for assessment among all the readings by selecting a period.


Click on the arrow symbol in the data fields to open a calendar, making selection easier.

Click on the month or year to change it immediately.

◀ Jun 2012 ▶

Mo	Di	Mi	Do	Fr	Sa	So
28	29	30	31	1	2	3
4	5	6	7	8	9	10
11	12	13	14	15	16	17
18	19	20	21	22	23	24
25	26	27	28	29	30	1
2	3	4	5	6	7	8

3.3 “Sys/dia profile” registry



This displays the blood pressure trend.

The pulse trend can be hidden or displayed by the Ctrl+P key combination.

Move the cursor over the measurement to obtain more information about individual readings. The date, time and reading are displayed in the top left-hand corner of the diagram.

3.3.1 “All, Morning, Afternoon” selection

You can restrict the reading selection to only morning measurements or only afternoon measurements here.

3.3.2 “Period” selection

☒ Period 06.04.2005 - 10.02.2012

You can select a specific period for assessment among all the readings by selecting a period.

Click on the arrow symbol in the data fields to open a calendar, making selection easier.

Click on the month or year to change it immediately.

◀ Jun 2012 ▶						
Mo	Di	Mi	Do	Fr	Sa	So
28	29	30	31	1	2	3
4	5	6	7	8	9	10
11	12	13	14	15	16	17
18	19	20	21	22	23	24
25	26	27	28	29	30	1
2	3	4	5	6	7	8

3.3.3 “Time scope” selection field

This allows you to select the period for which measurements are displayed. You have the choice of Auto, Day, Week, Month, Quarter or Year.

“Auto” is the default setting. In this setting, the time scope is scaled from the first measurement to the last.


Right-clicking in the diagram allows you to shift the graph horizontally with the mouse.

3.3.4 “Print...” button

This produces a print-out of the SYS/DIA profile, pulse profile and statistics.
If you selected “PDF file” in the printer settings (chapter 3.7.3), a PDF file will be created.
The name of this PDF file is medi_pat#_YYYYMMDD_sys.pdf

where
pat#=patient number
YYYYMMDD = date of measurement
sys = SYS/DIA reference letter

3.4 “MAP profile” registry



This displays the trend for the calculated mean arterial blood pressure. (MAP = diastole DIA + 1/3 pulse pressure PP).

The pulse trend can be hidden or displayed by the Ctrl+P key combination.

Move the cursor over the measurement to obtain more information about individual readings. The date, time and reading are displayed in the top left-hand corner of the diagram.

3.4.1 “All, Morning, Afternoon” selection

You can restrict the reading selection to only morning measurements or only afternoon measurements here.

3.4.2 “Period” selection

☒ Period 06.04.2005 - 10.02.2012

You can select a specific period for assessment among all the readings by selecting a period.

Click on the arrow symbol in the data fields to open a calendar, making selection easier.

Click on the month or year to change it immediately.



3.4.3 “Time scope” selection field

This allows you to select the period for which measurements are displayed. You have the choice of Auto, Day, Week, Month, Quarter or Year.

“Auto” is the default setting. In this setting, the time scope is scaled from the first measurement to the last.

Right-clicking in the diagram allows you to shift the graph horizontally with the mouse.


3.4.4 “Print...” button

This produces a print-out of the MAP profile, pulse profile and statistics.

If you selected “PDF file” in the printer settings (chapter 3.7.3), a PDF file will be created. The name of this PDF file is medi_pat#_YYYYMMDD_map.pdf

where
pat#=patient number
YYYYMMDD = date of measurement
map = MAP reference letter

3.5 “PP profile” registry



This displays the trend for the calculated pulse pressure. (PP = SYS-DIA).

The pulse trend can be hidden or displayed by the Ctrl+P key combination.

Move the cursor over the measurement to obtain more information about individual readings. The date, time and reading are displayed in the top left-hand corner of the diagram.

3.5.1 “All, Morning, Afternoon” selection

You can restrict the reading selection to only morning measurements or only afternoon measurements here.

3.5.2 “Period” selection

☒ Period 06.04.2005 ▼ - 10.02.2012 ▼

You can select a specific period for assessment among all the readings by selecting a period.

Click on the arrow symbol in the data fields to open a calendar, making selection easier.

Click on the month or year to change it immediately.

◀ Juni 2012 ▶

Mo	Di	Mi	Do	Fr	Sa	So
28	29	30	31	1	2	3
4	5	6	7	8	9	10
11	12	13	14	15	16	17
18	19	20	21	22	23	24
25	26	27	28	29	30	1
2	3	4	5	6	7	8

3.5.3 “Time scope” selection field

This allows you to select the period for which measurements are displayed. You have the choice of Auto, Day, Week, Month, Quarter or Year.

“Auto” is the default setting. In this setting, the time scope is scaled from the first measurement to the last.

Right-clicking in the diagram allows you to shift the graph horizontally with the mouse.


3.5.4 “Print...” button

This produces a print-out of the pulse pressure profile, pulse profile and statistics.

If you selected “PDF file” in the printer settings (chapter 3.7.3), a PDF file will be created. The name of this PDF file is medi_pat#_YYYYMMDD_pp.pdf

where
pat#=patient number
YYYYMMDD = date of measurement
pp = pulse pressure reference letter

3.6 “Statistics” registry



Measurements for the complete period, morning and afternoon periods are assessed separately.

The total number of readings in the individual periods, and the number of arrhythmias recognised, are displayed.

The following values are displayed in the individual columns:

Min: smallest value in the interval in question

Mean: arithmetic mean in the interval in question

Max: largest value in the interval in question

SD: standard deviation in the interval in question

> Limit: percentage above the limits that have been set

The percentage increase/decrease in the afternoon is displayed.

3.6.1 “Print...” button

This produces a print-out of the SYS/DIA profile, pulse profile and statistics.

If you selected “PDF file” in the printer settings (chapter 3.7.3), a PDF file will be created. The name of this PDF file is `medi_pat#_YYYYMMDD_sys.pdf`

where


pat#=patient number

YYYYMMDD = date of measurement

sys = SYS/DIA reference letter

3.7 “Settings...” button in “Patient” registry

3.7.1 “Doctor” sub-registry




The screenshot shows a 'Settings' dialog box with a tabbed interface. The 'Doctor' tab is selected, showing input fields for Name, Street, City, Telephone, and Fax. The other tabs are 'Import/Export', 'Printer Settings', 'Interface', 'Limits', and 'Profile'. The 'OK' and 'Cancel' buttons are at the bottom right.

Field	Value
Name	Dr.med. Mustermann
Street	Musterstraße 1
City	12345 Musterstadt
Telephone	01234 567 890
Fax	01234 567 891

The personal details of the doctor treating the patient are entered in the “Doctor” registry and used to produce the footer in the print-out.

3.7.2 "Import/Export" sub-registry



If you are using a clinic data processing system that supports the GDT interface, this is where the path and file names for export and import files are set. You have direct access to the Windows directory structure via the buttons next to the input fields for import and export files.

For example: c:\prax_ed\import.gdt
Import file = clinic DP system export file
Export file = clinic DP system import file

GDT import takes place automatically when you start the program, or manually by clicking the "Export/Import -> Import GDT" button if a valid GDT import file is present in the relevant directory.


GDT export takes place only if the patient being imported from the clinic DP system is active. Export takes place automatically when the program is closed down (if measurements have been read in from the device during the current program session) or can be performed manually at any time by clicking the "Export GDT" button in the "Measurements" registry.



Data export is deactivated if another patient has been selected after automatic import. It is reactivated when the first patient is reselected. Data export is also deactivated if no import has first taken place.

Further informations to the available options "Produce PDF file during GDT-export" and "GDT export permanently active" see 3.4.2 .

3.7.3 “Printer settings” sub-registry



In the “Printer settings” registry you decide whether the print-out should be produced in paper form, as a screen print and/or as a PDF document, once the relevant [Print...] buttons have been clicked while the program is running.


If you select “Printer”, then the “Output Options” dialogue appears when you click on the [Print...] buttons while the program is running. You can choose either a print output (“Printer”) or a screen view (“Preview”). The “File” function has no significance here.

If you select “PDF file”, a PDF file is produced. The name of this PDF file is created in different ways depending on the application, and this process is described in more detail in the individual applications.




You must select at least one medium: “Printer” or “PDF file”.

If you select the “monochrome” option, the print will be in black and white.
If you press the “Configured print-out” printer symbol, the print-out format configured here will be used and only the selected pages will be printed.



3.7.4 “Interface” sub-registry



If the “First available USB port” option is active, the first FTDI USB port under Windows will be used to transfer data. If more than one FTDI USB port is installed, disconnect the ports that are not needed from the computer, or deactivate the option.

If the “First available USB port” option is not active, the serial (COM) interface or USB port selected in the port list will be used for data transfer.

3.7.5 “Limits” sub-registry

Settings


DoctorImport/ExportPrinter SettingsInterfaceLimitsProfile

Total	Day	Night
Systole130	Systole135	Systole120
Diastole80	Diastole85	Diastole75
Pulse100	Pulse110	Pulse90
MAP100	MAP110	MAP90
PP50	PP55	PP45

OKCancel

The limits for the individual periods can be set in order to allow customised assessment. The limit settings for morning and afternoon are based on the morning and afternoon interval periods.

3.7.6 “Profile” sub-registry



In the profile displays (“Sys/Dia, MAP, PP profile” registry), an envelope curve (line connecting the individual measurements) is normally drawn, with the limits shown as a horizontal line.

Both the envelope curve and the limit lines can be hidden.

4. Warranty provisions

Irrespective of statutory warranty rights, boso guarantees the software data carrier to be free from material and processing defects for 90 days from the date of purchase.

In the light of the warranty undertaking given above, boso is obliged to replace the data carrier and the software it contains under the following conditions:

- the customer/purchaser notifies boso of the defect in writing within 90 days of the date of purchase

or

- the customer/purchaser returns the defective data carrier to the dealer, or direct to boso at the address below, within 90 days:

boso
BOSCH + SOHN GmbH u. Co. KG
Bahnhofstr. 64
72417 Jungingen
Germany

boso does not offer any warranty, either explicitly or tacitly, beyond the warranty undertaking given above.

5. Limitation of liability

- a) The software is supplied to you on the basis of the current state of development.
- b) Irrespective of the provision set out in chapter 4, bosco and its dealers accept no liability for risks with regard to the results and performance of the software. We bear no liability for indirect losses, consequential losses, loss of earnings, losses due to lost or damaged data or other commercial or financial losses.
- c) The aforementioned exclusion of liability does not apply to liability under product liability legislation or to cases of deliberate action or gross negligence on our part or on the part of our agents.

6. Explanation of icons



Read instruction manual



Year of manufacture



Manufacturer



Important advice

Appendix A - Extract from GDT record description

The record types and their field identifiers needed to create the GDT interface of this software are described below:

Record type 6302 "Request new examination"

Field 8000: record identification

8100: record length

8315: GDT ID of recipient

8316: GDT ID of sender

8410: Testident (BDM00: shortcut patient measurement)

9218: GDT version number

3000: patient number

3101: patient's surname

3102: patient's first name

3103: patient's date of birth

3105: insurance number

3106: patient's address (town)

3107: patient's address (street)

3110: patient's gender (1 = male, 2 = female)

3622: patient's height (cm)

3623: patient's weight (kg)

Record type 6310 "Transmit data from examination"

Field 8000: record identification

8100: record length

8315: GDT ID of recipient

8316: GDT ID of sender

9218: GDT version number

3000: patient number / patient identification

8402: Map for a particular device and process

6200: Date of examination

6228: Results table text, formatted

6302: File archiving identifier

6303: File format

6304: File content

6305: File reference



BOSCH + SOHN GmbH u. Co. KG
Bahnhofstraße 64 • 72417 Jungingen • Germany
Phone: +49 (0) 74 77 / 92 75-0 • Telefax: +49 (0) 74 77 / 10 21
www.boso.de • mail: zentrale@boso.de

Subject to errors and alterations.
(15/05/2014)



BOSCH + SOHN GmbH u. Co. KG
Bahnhofstraße 64 • 72417 Jungingen • Germany
Phone: +49 (0) 74 77 / 92 75-0 • Telefax: +49 (0) 74 77 / 10 21
www.boso.de • mail: zentrale@boso.de

Subject to errors and alterations.
(25/08/2014)